

# BREAKING THE RESOLUTION BARRIER IN OPTICAL MICROSCOPY: A NEW RESOLUTION MEASURE WITH APPLICATIONS TO SINGLE MOLECULE IMAGING

*Sripad Ram<sup>1,2</sup>, E. Sally Ward<sup>1</sup> and Raimund J. Ober<sup>\*1,3</sup>*

<sup>1</sup> Center for Immunology, University of Texas Southwestern Medical Center, Dallas, TX,

<sup>2</sup> Joint Biomedical Engineering Graduate Program, UT Southwestern Medical Center/UT Arlington,

<sup>3</sup> Department of Electrical Engineering, University of Texas at Dallas, Richardson, TX.

## ABSTRACT

Rayleigh's resolution criterion, although extensively used in optical microscopy, is well known to be based on heuristic notions. In fact, recent single molecule experiments have shown that this criterion can be surpassed in a regular optical microscope. The inadequacy of Rayleigh's criterion has necessitated a reassessment of the resolution limits of optical microscopes. Recently, we proposed a new resolution measure that overcomes the limitations of Rayleigh's criterion. Known as the fundamental resolution measure FREM, our new result predicts that distances well below Rayleigh's limit can be resolved in an optical microscope. The effect of deteriorating experimental factors on the new resolution measure is also investigated. Further, it is experimentally verified that distances well below Rayleigh's limit can be measured from images of closely spaced single molecules with an accuracy as predicted by the new resolution measure. We have also addressed an important problem in single molecule microscopy that concerns the accuracy with which the location of a single molecule can be determined. In particular, we have derived analytical expressions for the limit to the 2D/3D localization accuracy of a single molecule.

**Index Terms**— Fluorescence microscopy, 2D resolution, Rayleigh's resolution criterion, optical microscopes, Fisher information matrix, Cramer-Rao inequality, localization accuracy, parameter estimation

## 1. INTRODUCTION

The resolution of an optical system is an important parameter that quantifies the capability to discern two closely spaced point sources. In optical microscopy, Rayleigh's criterion is extensively used to determine the resolvability of microscopes. Despite its widespread use, it is well known that Rayleigh's criterion is based on heuristic notions. For instance, this criterion was formulated within a deterministic framework and therefore it does not take into account the

photon statistics of the acquired data. This is not surprising, since Rayleigh's criterion was developed at a time when the unaided human eye was typically used as a detector. Therefore, this criterion is not well adapted to modern imaging approaches that use highly sensitive photon counting cameras. In fact recent single molecule experiments have shown that Rayleigh's criterion can be surpassed in an optical microscope setup ([1, 2, 3, 4]), thereby illustrating that this criterion is inadequate for current microscopy techniques.

Recently, we proposed a new resolution measure that overcomes the limitations of Rayleigh's criterion and provides a quantitative measure of the microscope's ability to resolve two point sources ([4]). Known as the fundamental resolution measure FREM, the new result predicts that distances well below Rayleigh's criterion can be resolved in a conventional optical microscope setup ([4]). Here, we present a review of this and other related results that were recently reported by our group ([4, 5, 6, 8]). We adopted a general stochastic framework and made use of the statistical theory concerning the Fisher information matrix to derive an analytical formula for the new resolution measure. The new result predicts that the resolvability of the optical microscope can be improved by increasing the number of detected photons from the point sources. We also investigated the effect of deteriorating factors such as pixelation of the detector and extraneous noise sources on the new resolution measure ([4]). By imaging closely spaced single molecules and estimating their distance of separation, it was verified that distances well below Rayleigh's resolution limit can be resolved in an optical microscope setup with an accuracy as predicted by the new resolution measure ([4]).

In quantitative optical microscopy, parameter estimation approaches play a crucial role in the analysis of the acquired data (see [1, 3, 4, 5]). To carry out such analyses, it is helpful for the experimenter to have analytical tools to assess with which accuracy the various parameters can be estimated. The stochastic framework used to derive the new resolution measure can be adopted to address a variety of problems concerning quantitative data analysis in optical microscopy ([8]). As an application, we addressed an important problem in single

This research was supported in part by the National Institutes of Research (RO1 GM071048). \*Corresponding author, e-mail: ober@utdallas.edu.

molecule microscopy that concerns the accuracy with which the location of a single molecule can be determined ([5, 6]). In particular, we considered both in focus and out of focus scenarios and derived analytical expressions for the limit to the localization accuracy of a single molecule.

## 2. GENERAL RESULTS

In a typical quantitative imaging experiment, attributes of the object of interest such as location, distance of separation from other objects, orientation, photon count etc., are determined from the acquired data by using a specific estimation procedure. In any estimation problem, it is important to know whether the specific estimation technique used to estimate the desired attribute indeed comes close to the best possible accuracy. This can be determined by calculating the Fisher information matrix for the underlying parameter estimation problem. The Fisher information matrix  $\mathbf{I}(\theta)$  plays a central role in the theory of parameter estimation algorithms. Its inverse provides, through the classical Cramer-Rao inequality ([7]), a lower bound to the variance of any unbiased estimator  $\hat{\theta}$  of an unknown parameter  $\theta$ , i.e.,  $\text{var}(\hat{\theta}) \geq \mathbf{I}^{-1}(\theta)$ . Because the performance of estimation algorithms is typically specified in terms of its standard deviation, the above inequality implies that the square root (of the corresponding leading diagonal entry) of the inverse Fisher information matrix provides a lower bound to the performance of any unbiased estimator of  $\theta$ . Stated otherwise, the square root of the inverse Fisher information matrix provides a *limit* to the accuracy with which a specific object attribute can be determined.

### 2.1. Fisher information matrix

The data acquired in an optical microscope setup is modeled as a spatio-temporal random process, which we refer to as the image detection process  $\mathcal{G}$  ([8]). The temporal part is an inhomogeneous Poisson process that models the time points of the detected photons and the spatial part is a sequence of independent random variables that models the location coordinates at which the photons hit the detector. The general expression of the Fisher information matrix for the problem of estimating an unknown parameter  $\theta$  from the acquired data is given by ([8, 9])

$$\mathbf{I}(\theta) = \int_{t_0}^t \int_{\mathcal{C}} \frac{1}{\Lambda_{\theta}(\tau) f_{\theta, \tau}(r)} \left( \frac{\partial [\Lambda_{\theta}(\tau) f_{\theta, \tau}(r)]}{\partial \theta} \right)^T \times \frac{\partial [\Lambda_{\theta}(\tau) f_{\theta, \tau}(r)]}{\partial \theta} dr d\tau, \quad \theta \in \Theta. \quad (1)$$

In the above expression,  $\Lambda_{\theta}$  denotes the intensity function of the Poisson process,  $\{f_{\theta, \tau}\}_{\tau \geq t_0}$  denotes the density function of the independent random variables and  $\mathcal{C}$  denotes the detector. It is assumed that the spatial and the temporal components are mutually independent of each other. In deriving eq. 1 no

specific assumptions have been made regarding the functional form of  $f_{\theta, \tau}$  or  $\Lambda_{\theta}$ . Therefore, the above expression of  $\mathbf{I}(\theta)$  is applicable to a wide variety of imaging conditions, such as (in)coherent/polarized illumination and detection, etc. We note that an implication of the time dependence of the density function  $f_{\theta, \tau}$  is that the above equation is applicable to moving objects.

## 3. APPLICATIONS

### 3.1. Resolution beyond Rayleigh's criterion

The advent of single molecule microscopy has generated significant interest in studying nano-scale interactions within a cellular environment. It is widely believed that Rayleigh's criterion impedes the study of single molecular interactions at distances below 200 nm. Fluorescence resonance energy transfer based techniques are typically used to probe biomolecular interactions up to a distance of 10 nm. This, however, leaves a gap in the distance range of 10 - 200 nm which is important for the study of many biological processes in an optical microscope. It has been suggested that Rayleigh's resolution limit can be overcome when apriori information in conjunction with parameter estimation approaches are used to analyze the acquired data ([10]). In fact by using such approaches, several groups have shown that Rayleigh's limit can be surpassed in an optical microscope setup ([1, 2, 3, 4]).

By using the methodology based on the Fisher information matrix laid out above we have proposed a new resolution measure that overcomes the limitations of Rayleigh's criterion ([4]). Known as the fundamental resolution measure FREM, the new resolution measure predicts that distances well below Rayleigh's limit can be resolved in an optical microscope. An analytical formula for the FREM has been derived and is given by ([4])

$$\delta_d := \frac{1}{\sqrt{4\pi \cdot \Lambda_0 \cdot (t - t_0) \cdot \Gamma_0(d)}} \cdot \frac{\lambda}{n_a}, \quad (2)$$

where  $\lambda$  denotes the emission wavelength of the detected photons,  $n_a$  denotes the numerical aperture of the objective lens,  $\Lambda_0$  denotes the photon detection rate (intensity) per point source,  $[t_0, t]$  denotes the acquisition time interval, and  $\Gamma_0(d)$  is given by

$$\Gamma_0(d) := \int_{\mathbb{R}^2} \frac{1}{\frac{J_1^2(\alpha r_{01})}{r_{01}^2} + \frac{J_1^2(\alpha r_{02})}{r_{02}^2}} \left( \left(x + \frac{d}{2}\right) \frac{J_1(\alpha r_{01}) J_2(\alpha r_{01})}{r_{01}^3} - \left(x - \frac{d}{2}\right) \frac{J_1(\alpha r_{02}) J_2(\alpha r_{02})}{r_{02}^3} \right)^2 dx dy,$$

with  $J_n$  denoting the  $n^{th}$  order Bessel function of the first kind,  $\alpha := 2\pi n_a / \lambda$ ,  $r_{01} := \sqrt{(x + d/2)^2 + y^2}$  and  $r_{02} := \sqrt{(x - d/2)^2 + y^2}$ . From eq. 2 we see that the resolution measure is given in terms of quantities such as the expected number of detected photons, numerical aperture of the objective lens, and wavelength of the detected photons. The

above expression was derived for imaging conditions analogous to those of Rayleigh's criterion, i.e., two equal intensity, in-focus point sources that emit incoherent, unpolarized light. In many applications, however, these conditions are not satisfied. Hence a general analytical expression for the resolution measure has been derived that is applicable to a variety of imaging conditions ([4]). The FREM provides the best-case scenario for a microscope setup, where experimental factors that potentially deteriorate the acquired data are not taken into account. This was done to obtain an expression for the best possible resolution in the absence of deteriorating factors due to specific experimental settings. Current imaging detectors have pixels and therefore the acquired data is a discretized version of the original image. Aside from this, the acquired data contains extraneous noise sources. Here two independent, additive noise sources are considered, i.e., Poisson noise and Gaussian noise. Poisson noise can be used to model, for example, the effect of scattered photons on the measured data and Gaussian noise characterizes, for example, measurement noise in the CCD detector ([11]). Analytical expressions for the resolution measure that take into account these deteriorating factors have also been obtained ([4]). The latter result is referred to as the practical resolution measure PREM.

By definition, the new resolution measure is a bound to the accuracy with which the distance between two point sources can be resolved. To verify if the resolution measure can be attained in a practical experimental setup, images of closely spaced single molecules were acquired and their distance of separation were estimated by using the maximum likelihood estimator (see [4] for details). It was found that distances well below Rayleigh's resolution limit can be determined from the acquired data with an accuracy as predicted by the new resolution measure.

### 3.2. Single molecule localization accuracy

One of the central problems in single molecule data analysis concerns the accuracy with which the location of a single molecule can be determined. By using the methodology based on the Fisher information matrix, we have investigated the single molecule localization accuracy problem. We derived a simple analytical formula that provides the fundamental limit to the accuracy with which the 2D location coordinates  $(x_0, y_0)$  of a single molecule can be determined ([5, 8]), which is given by

$$\delta_{x_0}^{2d} = \delta_{y_0}^{2d} = \frac{\lambda}{2\pi n_a \sqrt{\Lambda_0(t - t_0)}}, \quad (3)$$

where  $\lambda$  denotes the wavelength of the detected photons,  $n_a$  denotes the numerical aperture of the objective lens,  $\Lambda_0$  denotes the photon detection rate of the single molecule and  $[t_0, t]$  denotes the acquisition time interval. The importance of this result lies in the fact that it shows with an unexpectedly simple expression how fundamental properties of the single

molecule (emission wavelength, photon-detection rate) and of the detection system (numerical aperture, acquisition time) influence the localization accuracy of the single molecule. The above result is referred as 'fundamental', since the model that underlies the derivation of the result does not take into account deteriorating experimental factors such as pixelation of the detector or extraneous noise sources in the acquired data. Therefore, the above result pertains to the best-case scenario for a given imaging configuration.

Eq. 3 provides an expression for the 2D fundamental limit to the localization accuracy of a single molecule, where it is assumed that the single molecule lies in the focal plane of the objective lens (in-focus scenario). However, in a cellular environment a single molecule can move in all three dimensions. Therefore, for such cases it is important to know the 3D limit of the localization accuracy. In [6] we have addressed this problem and have obtained analytical expressions for the 3D fundamental limit to the localization accuracy of a single molecule.

The derivation of the fundamental limit given in eq. 3 assumes the best case scenario for the acquisition system. This was done to obtain an expression for the best possible localization accuracy in the absence of deteriorating factors due to specific experimental settings. Analytical expressions for the 2D and the 3D limit of the localization accuracy have been obtained that take into account deteriorating factors ([5, 6]).

It should be pointed out that the stochastic framework described in Section 2 allows for both stationary and time varying image profiles. Therefore analytical expressions for the limit of the localization accuracy of moving objects can also be obtained.

### 3.3. Multifocal plane single molecule imaging

One of the shortcomings of conventional wide-field optical microscopes is their poor depth discrimination capability. Due to this, there exists significant uncertainty in determining the axial location of point objects (e.g., single molecules), especially when they are close to the plane of focus. Previously we showed that the limit of the 3D localization accuracy of a single molecule can significantly vary depending upon the defocus level ([6]). For instance, for small defocus values ( $\leq 200$  nm), it was predicted that the  $x_0/y_0$  coordinate of the single molecule can be determined with relatively high accuracy whereas the  $z_0$  coordinate can be determined with poor accuracy. On the other hand, for large defocus values (200 - 700 nm), it was predicted that the  $x_0/y_0$  coordinate of the single molecule can be determined with poor accuracy whereas the  $z_0$  coordinate can be determined with high accuracy. Due to this mismatch in the limit of the localization accuracy between the  $x_0/y_0$  coordinate and the  $z_0$  coordinate, it is difficult to determine all three coordinates with the same level of accuracy.

To overcome this problem, we propose to use the mul-

tifocal plane imaging technique that was previously developed by our group ([12, 13]). Here we simultaneously image two distinct focal planes within the specimen. For example, one of the focal planes could correspond to the standard focal plane that is imaged in a conventional widefield microscope, while the other focal plane could correspond to a plane that is shifted away from the standard focal plane. If single molecules are imaged in the above setup, then the image of the shifted focal plane provides additional information regarding the single molecule location. By making use of this additional data, we expect to obtain higher accuracy in determining the location of the single molecule. An alternative to the multifocal plane imaging technique is to use a focusing device, which *sequentially* moves the objective lens to acquire images of the different focal planes. A drawback of this approach is that focusing devices are typically slow and moreover, suffer from the lack of synchrony between their movement and the movement of the single molecules in the specimen. Thus when the specimen is being imaged at one focal plane important events can be missed in the other planes. With the multifocal plane imaging approach these problems are avoided, since there is no movement of the objective lens and more importantly the specimen is simultaneously imaged at multiple planes.

#### 4. REFERENCES

- [1] M. P. Gordon, T. Ha, and P. R. Selvin, "Single molecule high resolution imaging with photobleaching," *Proceedings of the National Academy of Sciences USA*, vol. 101, pp. 6462–6465, 2004.
- [2] X. Qu, D. Wu, L. Mets, and N. F. Scherer, "Nanometer-localized multiple single-molecule fluorescence microscopy," *Proceedings of the National Academy of Sciences USA*, vol. 101, no. 31, pp. 11298–11303, 2004.
- [3] K. A. Lidke, B. Rieger, T. M. Jovin, and R. Heintzmann, "Superresolution by localization of quantum dots using blinking statistics," *Optics Express*, vol. 13, no. 18, pp. 7052–7062, 2005.
- [4] S. Ram, E. S. Ward, and R. J. Ober, "Beyond Rayleigh's criterion: A resolution measure with application to single-molecule microscopy," *Proceedings of the National Academy of Sciences USA*, vol. 103, pp. 4457–4462, 2006.
- [5] R. J. Ober, S. Ram, and E. S. Ward, "Localization accuracy in single molecule microscopy," *Biophysical Journal*, vol. 86, pp. 1185–1200, 2004.
- [6] S. Ram, E. S. Ward, and R. J. Ober, "How accurately can a single molecule be localized in three dimensions using a fluorescence microscope?," *Proceedings of the SPIE*, vol. 5699, pp. 426–435, 2005.
- [7] C. R. Rao, *Linear statistical inference and its applications*, Wiley, New York, USA., 1965.
- [8] S. Ram, E. S. Ward, and R. J. Ober, "A stochastic analysis of performance limits for optical microscopes," *Multidimensional Systems and Signal Processing*, vol. 17, pp. 27–58, 2006.
- [9] D. L. Snyder and M. I. Miller, *Random point processes in time and space*, Springer Verlag, New York, USA, 2nd edition, 1991.
- [10] P. C. D. Hobbs, *Building electro-optical systems*, John Wiley and Sons, 2000.
- [11] G. C. Holst, *CCD arrays, cameras and displays*, SPIE Press, Bellingham, WA, 2nd edition, 1998.
- [12] P. Prabhat, S. Ram, E. S. Ward, and R. J. Ober, "Simultaneous imaging of different focal planes in fluorescence microscopy for the study of cellular dynamics in three dimensions," *IEEE Transactions on Nanobioscience*, vol. 3, no. 4, pp. 237–242, 2004.
- [13] P. Prabhat, S. Ram, E. S. Ward, and R. J. Ober, "Simultaneous imaging of several focal planes in fluorescence microscopy for the study of cellular dynamics in 3D," *Proceedings of the SPIE*, vol. 6090, pp. 115–121, 2006.